

Table 1 Atomization conditions for hair lead analysis

Step	Temp (°C)	Ramp (s)	Hold time (min)	Argon flow (mL)
Desiccation	150	15	20	300
Combustion	1100	15	10	300
Atomization (reading)	1600	0	3	0
Cleaning	2500	1	3	300

Temp = temperature.