

**Table 2 Maximum levels allowed for chemical contaminants, to be tested once per year**

<b>Chemical contaminant</b>	<b>Maximum recommended concentration (mg/l=ppm)</b>
Lead	0.005
Aluminum	0.01
Nitrate (as N)	2
Total chlorine	0.1
Copper	0.1
Fluoride	0.2
Magnesium	4 (0.15 mmol/L)
Calcium	2 (0.05 mmol/L)
Potassium	8 (0.2 mmol/L)
Sodium	70 (3 mmol/L)