

Table 3 Maximum allowable levels for TDS, Chlorine, Total viable microbial count (TVC) and Endotoxins, to be tested once per month

Parameter tested	Maximum allowable level in ultrapure water
TDS	10 ppm
Total Chlorine	0.1 mg/l
TVC	0.1 CFU/mL
Endotoxins	0.03 EU/mL