

Table 1 Accuracy of the results of susceptibility testing of 291 laboratories sent Staphylococcus saprophyticus

Antibiotic	Expected	Correct No. of labs	Incorrect No. of labs	No answer No. of labs
Ampicillin	S	70	172	4
Tetracycline	S	254	27	10
Nalidixic acid	R	273	13	5
Ciprofloxacin	S	273	12	6
Co-trimoxazole	S	239	43	9
Total		1109 (89.3%)	267 (8.2%)	34 (2.6%)

R = resistant, S = susceptible, I = intermediate.